



Recently BAPI changed its certification form to match the requirements of the National Institute of Standards and Technology data reporting standard.

<b>BAPI</b>		<b>CERTIFICATE OF CALIBRATION</b>					
Customer	Your Company Name		Order #	Your Order #	CalDate	2/17/2006	
Serial #	BCC146	BA/T1KM[-40 TO 120F]-O-WP			CalDue	5/18/2006	
Certificate #	BCEC1226	Cal. Procedure	T1KCalibration.pdf		Calibrated By	Tim VanBlarcom	

The first section (shown above) indicates the product being certified, the customer and the order number.

<b>Environmental Conditions</b>							
Humidity %RH	38	Temperature °F	71.6	Pressure	1016	Pascals	

The second section (shown above) records the relative humidity, temperature and atmospheric pressure of the test laboratory.

<b>Calibration Standards</b>							
BAPI ID#	Description			Uncertainty			
BAPI0016	SPRT			.02°C			
BAPI0015	Digital Thermometer			.02°C			
BAPI0116	Digital Multimeter			.001%			

The third section (shown above) is an inventory of the equipment used to perform the certification. Uncertainty is the tolerance of the instrument's measurement as determined during its last calibration at a NIST certified calibration center.

<b>Results</b>							
Test	Units	Reference	Uncertainty	As Found	Difference	As Left	Difference
00.0	°F	-0.2	.04°C	00.0	.2	-0.2	0
40.0	°F	40.0	.04°C	40.3	.3	40.1	.1
80.0	°F	80.6	.04°C	80.8	.2	80.6	0
Notes							

The fourth section (shown above) details the certification results. The column labeled Test defines the test procedure or procedures that were used to certify the product. The column labeled Units defines the units of measure used for the test. The three test conditions for this certification were 0°F, 40°F and 80°F.





The column labeled Reference is the actual test condition as measured by the Calibration Standards referenced in the inventory described above. For the test shown, the test condition at 0°F was actually -0.2°F, we achieved 40°F and 80°F was actually 80.6°F. The next column labeled Uncertainty is the tolerance of Calibration standards used to measure the test condition temperature.

The column labeled As Found is the transmitters output before any corrections are made to the transmitter. If the output is a 4 to 20mA current loop, the output is changed to the units of the parameter being measured, in this case Fahrenheit temperature. The next column labeled Difference is the difference or offset of the As Found to the Reference.

The column labeled As Left is the transmitters output after any corrections are made to the transmitter. This is how the equipment is sent to you. If the output is a 4 to 20mA current loop, the output is changed to the units of the parameter being measured, in this case Fahrenheit temperature. The next column labeled Difference is the difference or offset of the As Left to the Reference. This last Difference column is the offset you should use in your controller to correct the temperature.

**This notification serves to certify that the unit described above has been inspected and tested in accordance with specifications published by Building Automation Products Inc.**

**The accuracy and calibration of this instrument are traceable through reference standards that are compared, at planned intervals, to national standards maintained by the National Institute of Standards and Technology (NIST), by comparison to natural physical constants.**

**The measurement standards which support this calibration are calibrated on a schedule to maintain the required accuracy level.**

For recalibration and recertification of this unit or for other testing or calibration services contact:



Building Automation Products, Inc.  
750 North Royal Avenue  
Gays Mills, WI 54631

Phone (608) 735-4800  
Fax (608) 735-4804

The last section (shown above) notifies you that all instruments used to certify the equipment are properly calibrated and traceable to NIST.

Additional information on specific Temperature, Pressure and Humidity Certification documents is found on the following three pages.

If you have any questions about the certification documents, please contact your BAPI representative.

